

**OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER: BARGARH
(HEALTH & F.W. DEPTT., GOVT. OF ORISSA)
TENDER CALL NOTICE**

E-mail id substore.bargarh@gmail.com Tel / Fax: 06646 -232804

TENDER FOR EQUIPMENTS, INSTRUMENTS AND CONSUMABLES UNDER NOHP, BARGARH

Bid Reference No. – 6484/Bargarh/dt.18.08.2025

DATE OF COMMENCEMENT & DOWNLODING OF THE BID DOCUMENT:Dt.19.08.2025

LAST DATE & TIME OF RECEIPT OF BID DOCUMENTS: Dt. 08.09.2025,Upto 4 P.M

DATE & TIME OF PRE-BID DISCUSSION:- Dt. 26.08.2025,dt.11 A.M

DATE & TIME OF OPENING OF TECHNICAL BID: Dt.10.09.2025,Time-11 A.M

DATE & TIME OF OPENING OF FINANCIAL BID: Dt.10.09.2025,Time-03 P.M

PLACE OF OPENING OF BID DOCUMENTS: DTU, NHM, BARGARH

**ADDRESS FOR COMMUNICATION
AND
RECEIPT OF BID DOCUMENTS**



**O/o. C.D.M & P.H.O, BARGARH
BARGARH – 768028**

SALE OF TENDER / BID DOCUMENT

The bidders have to download the Tender Documents directly from the WEBSITE available at <https://bargarh.odisha.gov.in/>. The Tender processing **cost fee of Rs.1,000/- + 18% GST (one thousand + 18 % GST) & EMD amount of Rs.30,000/-** (Rupees thirty thousand only) **in shape of separate** Demand Drafts drawn in favor of ZSS, NON-NRHM FUNDS, payable at Bargarh should be enclosed along-with the Bid documents in Cover-A, i.e technical bid & sealed price bid should be enclosed in another envelope Cover-B, i.e price bid. Then both the envelopes should be enclosed in the third envelope. The Bidders should specifically superscribe, “**TENDER FOR EQUIPMENTS, INSTRUMENTS AND CONSUMABELS UNDER NOHP**” on the top left corner of the outer envelope containing Bid.

The tender paper will be rejected if the bidder changes any clause or Annexure of the bid document downloaded from the website.

TERMS AND CONDITIONS:-

- 1.1 Sealed tenders will be received by dated 08.09.2025 upto **4 P.M** by the CDM & PHO, Bargarh through Courier/Speed Post/ Regd. Post only for the purchase of **equipments, instruments and consumable items under NOHP**. Any tender received after the due date & time will be rejected / returned to the sender unopened.
- 1.2 The bidder(s) are to submit their tender in sealed covered envelopes for technical bid and financial bid which should be super scribed as **“TENDER FOR EQUIPMENTS, INSTRUMENTS AND CONSUMABLES UNDER NOHP”**.
- 1.3 The Sealed tenders submitted by the bidders will be opened at the DTU,**O/o CDM & PHO , Bargarh** on **dt.10.09.2025, at 11.00 AM**. The bidder or their only duly authorized representative is allowed to be present during the opening of the tenders if they so like.
- 1.4 No tender document(s) can be accepted after the expiry of scheduled date and time for receipt of bids.
- 1.5 The details of items and specification are mentioned at **Annexure – I**. The quantity to be procured may vary as per the demand / need.
- 1.6 Rate should be quoted in Indian Currency (with paisa in two decimals only), both in words and figures against each item as the payments will be made in Indian currencies only (as per **Annexure-II**). The bidder shall not quote the rate for any item other than the item specified in the list.
- 1.7 The GST will be charged as per the guidelines given by the Finance Dept., Govt. of Odisha from time to time.
- 1.8 Each page of the bid document shall be duly numbered, signed and self attested by the bidder.

A. Tender Processing Fee and EMD:

The bidder should furnish the EMD of **Rs.30,000/- (Thirty thousand only)** and Rs.1,000.00+18% GST towards processing fee (non-refundable) in shape of Bank Draft in favor of **“ZSS, NON-NRHM FUNDS payable at Bargarh”**. However the EMD of the non-responsive bidders will be returned after completion/finalization of the tender process and the EMD of the approved bidders will be kept as performance security for one year and return back after 1 year without interest.

B. ELIGIBILITY CRITERIA

Wholesalers / distributors / manufacturers are eligible to participate in the tender provided, they have:

- (i) Valid Manufacturing / Wholesale / Distributor License from the Drug Controller / competent authority.
- (ii) Proof of supply to any Govt. hospital / PSU hospital / corporate sector hospital
- (iii) Valid up-to-date TAX clearance certificate (Income Tax & GST).
- (iv) Annual Turnover of Rs. 10 lakhs or more in each year in last three preceding years
- (v) Shall have PAN Card, GST registration.
- (vi) Bidders are requested not to quote of products of those manufacturing unit who has been blacklisted either by the tender inviting authority or by any state Govt. or Central Govt. organization.

C. DOCUMENTS SHOULD BE SUBMITTED WITH THE TECHNICAL BID (COVER-A):-

1. Tender processing fee of Rs.1,000/- + 18% GST.
2. Earnest Money Deposit (EMD) of Rs.30,000/-.
3. **Copy of latest I.T Return & GST payment receipt.**
4. **Copy of PAN CARD & GST registration certificate.**
5. Copy of Valid manufacturing / wholesale / Distributor license from the Drug Controller / competent authority.
6. List of items quoted (**Annexure – I(A)**)
7. Authorization in original of the manufacturer in the name of the bidder (Annexure-V) along with product certification copies.
8. Price Bid in hard copy **signed & sealed** in each page properly along with the soft copy of the quoted price.
9. Details name, address, telephone no., Fax, e-mail of the bidder.
10. Declaration as per format Annexure-III.
11. Annual Turnover as per format Annexure – IV

D. PRICE BID

1. The tender format giving the quoted rate for the items should be sent in a separate sealed covers hereafter called **Cover “B” (Price Bid)**.
2. The tender format (Price Schedule) in duplicate in the prescribed form (as per **Annexure – II**), both hard copy and soft copy must be submitted in the bid. The price of the item should be quoted inclusive of insurance, packing, forwarding, freight (door delivery) but exclusive

of Taxes . The rate should be quoted for each item both in figures and words. Only two decimal points (paise) will be taken into consideration ignoring the rest digits. **In case of difference in words and figures, words will be taken into consideration for evaluation.**

4. The quoted rates should be final and shall not be subject to any escalation during the validity of the tender.
5. The bidders should submit/furnish a certificate in the bid to the effect that the price quoted by them is not more than the open market price or also under GeM Rate Contract where such rates exist. However, in circumstances when the price decreases during the contract period, the approved supplier ethically, should intimate the same to the purchasing authority.

E. REJECTION OF THE TENDER:

The tender paper will be rejected, if any of the following documents are wanting /not found with the tender bid:

- (i) Non submission of relevant documents as mentioned in Clause-C.
- (ii) Price bid / quoted rate with signature and seal.

F. EVALUATION:

1. Bidders will be evaluated as per the requirement of the bid and the price bid will be opened only for the bidders who qualified in the technical evaluation.
2. The cost of the unit item excluding Taxes / GST will be evaluated.
3. If the approved lowest eligible supplier fails to supply items in time, to meet the requirement the same shall be procured from the L2/ L3 suppliers at L1 rate, if they agree to supply at L1 approved rate.

G. Performance Security:

1. The EMD of the approved supplier(s) will be kept as performance security for one year. The EMD will be returned after 1 year without interest.

H. SUPPLY & DELIVERY:

1. Supply shall be made within 30 days from the date of issue of purchase order or as mentioned in the purchase order.
2. Delivery shall be made at Sub-Store, O/o CDM & PHO, Bargarh.
3. Non-supply / unwilling to supply the ordered quantity will lead to forfeiture of EMD submitted by the supplier. The defaulter supplier will be black listed for two years and the bidder will not participate in any tender published by the undersigned.

I. **PAYMENT:**

1. 100 % payment shall be made after submission of stock entry certificate(s) from the concerned authority and as per the availability of funds. Under no circumstances the supply should be interrupted as regards to payment.
2. **No advance payments towards cost of items supplied / turnkey job will be made to the approved bidder.**

J. **PENALTIES:**

1. If any item after use found to be **“Not of Satisfactory Quality”/Not as per the parameter/ Specification**”, such item will be declared as “Not of Satisfactory Quality” on the basis of the report of the concerned user. The said product shall be frozen. The supplier has to replace fresh stock as per the purchased / supplied quantity and take back the frozen stock. In case the supplier fails to replace the stocks, the EMD will be forfeited. No further purchase order will be placed to the firm / supplier for the item(s) and the firm / supplier will be blacklisted/debarred from participating in any tender floated in future for two years.

K. **GENERAL CONDITIONS:**

1. The tender documents should be clearly written /typed without any correction, interpolations and overwriting. Each page of the tender should bear the dated signature of the bidder.
2. All copies of the tender document should be self-attested. If any information or documents furnished by the bidder found to be misleading/incorrect at any stage, their tender will not be accepted.
3. The approved rate and supplier will be valid for **one year** from the date of approval or till finalization of the next tender whichever is later.
4. If the last date of submission of bid being declared as a holiday for the purchaser’s office, the due date submission of bids and opening of bids will be the following working date & time.
5. The quantity of procurement will be made on requirement basis.
6. The authority reserves the right to accept /reject all the bids or any part thereof without assigning any reason thereof.
7. All legal disputes, if any relating to purchase etc. are subject to jurisdiction in the courts of law situated at Bargarh, Odisha or High Court of Odisha.


Chief District Medical & Public Health Officer,
Bargarh

COMMON PRODUCT QUALITY STANDARD FOR ALL SURGICAL & DENTAL INSTRUMENTS:

1. Product should be BIS or CE or USFDA certified/registered. CE certificate must be issued by European authority in compliance to medical device directive (93/42/EEC) for surgical instruments and USFDA certificate/ registration must be supported by valid certificates.
2. Manufacturer should be WHO GMP certified.
3. Manufacturer should have ISO 13485 standard.
4. The instrument steel should comply to EN ISO 7153 standard for stainless steel used for surgical instruments and ISO 13402 standards: Surgical and dental hand instruments- Determination of resistance against autoclaving, corrosion and thermal exposure. Necessary certificate should be furnished. (Should be supported by test report or certificate from competent authority).
5. The instrument steel should comply to EN ISO 7153 standard for stainless steel used for surgical instruments and ISO 13402 standards: Surgical and dental hand instruments- Determination of resistance against autoclaving, corrosion and thermal exposure. Necessary certificate should be furnished. (Should be supported by test report or certificate from competent authority).
6. The Instruments must be vacuum-hardened in order to receive the necessary hardness tenacity and corrosion resistance. Detail must be furnished.
7. Stainless steel material Grade should be AISI 410 and 420 certified and should be supported by test reports.
8. Material should be Stainless Steel (Instruments) and Tungsten Carbide (TC) as mentioned.
9. Stainless steel material Grade should be AISI 410 and 420 certified and should be supported by test reports.
10. The Instruments should have a certificate of Primary passivation process.
11. Instruments should carry Laser marking indicating the Manufacturer's name and Batch No./ Lot No./ bar code, CE mark and Country of Origin
12. The inserts of Tungsten carbide instruments should be Laser welded not soldered.
13. All instruments should be provided with 3yrs. of replacement warranty.

List of Consumables Items under NOHP

1	Dental Burs round Small	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
2	Dental Burs round Medium	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
3	Dental Burs Endo access bur	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
4	Dental Burs Straight Fissure (SF21)	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
5	Dental Burs Straight Fissure (SF31)	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
6	Dental Burs Straight Fissure (SF41)	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
7	Dental Burs Inverted cone Small	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
8	Dental Burs Inverted cone Medium	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
9	Dental Burs Tapered Fissure (TF12)	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
10	Dental Burs Tapered Fissure (TF13)	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each

		mentioned instrument product quality standards.	
11	Glass Ionomer cement (Type-I)(Powder and liquid)	The cement packet should consist of powder and liquid in a separate bottle with powder dispensing scoop, paper pad and mixing spatula. There should be at least 10ml of liquid and 30gm of powder in the packet. The expiry date should be at least 1 year from the date of supply.	Packet
12	Glass Ionomer cement (Type-II) (Powder and liquid)	The cement packet should consist of powder and liquid in a separate bottle with, powder dispensing scoop, paper pad and mixing spatula. There should be at least 10ml of liquid and 30 gm of powder in the packet. The expiry date should be at least 1 year from the date of supply.	Packet
13	Matrix retainer (No. 8)	The retainer should be made up of high quality stainless steel. It should have long handle with pointed tips and auto cleavable.	Piece
14	Matrix retainer (No. 1)	The retainer should be made up of high quality stainless steel and auto-cleavable.	Piece
15	Tofflemire Retainer	The tofflemire matrix retainer should be of universal size and made up of high strength stainless steel. It should be autoclavable.	Piece
16	Matrix band (No. 8)	It should be made up of high quality stainless steel. Each packet should contain 12 no of matrix bands of different sizes.	Packet
17	Matrix band (No.1)	It should be made up of high quality stainless steel. Each packet should contain 12 no of matrix bands.	Packet
18	Tofflemire matrix band	The system should consist of matrix retainer and matrix band (30-40) for premolars and molars in small, medium and large size.	Packet
19	Zinc oxide powder	The pack should contain zinc oxide powder of 100mg in a plastic bottle. The expiry date should be at least for 1 year from the date of supply.	Packet
20	Eugenol	The pack should contain eugenol solution in 10ml bottles. The expiry date should be at least for 1 year	Vail/Packet

		from the date of supply.	
21	Composite restorative material for tooth color restoration-A1	The packet should contain a 3.5gm syringe of composite filling material of shade A1.The expiry date should be at least for 2 years from the date of supply.	Packet
22	Composite restorative material for tooth color restoration-A2	The packet should contain a 3.5gm syringe of composite filling material of shade A2.The expiry date should be at least for 2 years from the date of supply.	Packet
23	Composite restorative material for tooth color restoration-A3	The packet should contain a 3.5gm syringe of composite filling material of shade A3.The expiry date should be at least for 2 years from the date of supply.	Packet
24	Composite restorative material for tooth color restoration-B1	The packet should contain a 3.5gm syringe of composite filling material of shade B1.The expiry date should be at least for 2 years from the date of supply.	Packet
25	Composite restorative material for tooth color restoration-B2	The packet should contain a 3.5gm syringe of composite filling material of shade B2.The expiry date should be at least for 2 years from the date of supply.	Packet
26	Composite restorative material for tooth color restoration-Universal Shade	The packet should contain a 3.5gm syringe of composite filling material of shade-Universal.The expiry date should be at least for 2 years from the date of supply.	Packet
27	Universal Bonding Agent	The dentin bonding agent should be available in a bottle Containing 5gm of bonding agent.	Packet
28	Applicator tip	The packet should contain atleast 100 applicator tips in each pack.	Packet
29	37% etchant gel	The bottle should contain 37% phosphoric etchant liquid-gel. The bottle should contain atleast 9ml of etchant gel.The expiry date should be at least for 1 year from the date of supply.	Packet
30	Flowable composite	The packet should contain a 2gm syringe of flowable composite filling material of shade A1.The expiry date should be at least for 1 year from the date of supply.	1 Unit
31	Flowable composite	The packet should contain a 2gm syringe of flowable composite filling material of shade A2.The expiry date should be at least for 1 year from the date of supply.	1 Unit
32	Alginate impression material	The packet should contain 450gm of dust free powder. It should have good flavor. The expiry date should be at least for 1 year from the date of supply.	Packet
33	Dental Stone plaster	The packet should contain high quality dental stone plaster / green stone packed 3kg.	Packet
34	Calcium Hydroxide paste	The pack should have one tube of radio-opaque calcium hydroxide paste. The expiry date should be at least for 1 year from the date of supply.(2.5 ml / syringe, Disposable tips & Radio-opaque	Packet
35	EDTA gel for Root Canal	The packet should contain combination of ethylene diametetracetic acid(EDTA) and carbamide peroxide in water soluble gel base form of 5g syringe (pack of 3) with screw design stopper. The packet should contain at least 15 applicator tips. The expiry date should be at least for 1 year from the	Packet

		date of supply.	
36	K-file (10 no)(21mm)	The assorted pack of K-file (10)(21mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
37	K-file (10 No)(25mm)	The assorted pack of K-file (10)(25mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
38	K-file (10 no)(28mm)	The assorted pack of K-file (10)(28mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
39	K-file (15-40 no)(21mm)	The assorted pack of K-file (15-40)(21mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
40	K-file (15-40 no)(25mm)	The assorted pack of K-file (15-40)(25mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
41	K-file (15-40 no)(28mm)	The assorted pack of K-file (15-40)(28mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
42	K-file (45-80) (21mm)	The assorted pack of K-file (45-80)(21mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
43	K-file (45-80) (25mm)	The assorted pack of K-file (45-80)(25mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
44	H-file (15-40)(21mm)	The assorted pack of H-file (15-40)(21mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
45	H-file (15-40)(25mm)	The assorted pack of H-file (15-40)(25mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
46	H-file (45-80)(21mm)	The assorted pack of H-file (45-80)(21mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
47	H-file (45-80)(25mm)	The assorted pack of H-file (45-80)(25mm) should have good flexibility, high quality cutting efficiency and fracture resistance.	Packet
48	Sodium Hypochlorite 3%	500 ML Bottle	Bottle
49	2% GP Point	Size 15-100	1 Unit
50	2% GP Point(-2% (15-40)(25mm)	The assorted pack of GuttaPercha points-2% (15-40)(25mm) should be of high quality.	1 Unit
51	4% GP Point (15-40)(25mm)	The assorted pack of GuttaPercha points-4% (15-40)(25mm) should be of high quality.	Packet
52	2% GP Point (45-80)(25mm)	The assorted pack of GuttaPercha points-2% (45-80)(25mm) should be of high quality	Packet
53	2% Paper Points (15-100)	The 25-2% paper point should be of high quality for various endodontic uses. The expiry date should be at least for 2 years from the date of supply.	Packet

54	Calcium Hydroxide based root canal sealer.	The packet should contain base tube (15gm) and catalyst tube (15gm) with mixing pad. The expiry date should be at least 1 year from the date of supply.	Packet
55	Fluoride Varnish	The packet should contain a syringe/clicker of 10 mg. The expiry date should be at least for 1 year from the date of supply.	Packet
56	Pit and fissure sealant	The material should work with wet-bonding resin technology in a 3ml syringe pack. The expiry date should be at least for 1 year from the date of supply.	Packet
57	Plastic Spatula	Plastic Base	Packet
58	Steel Spatula	Stainless Steel Plastic Handle	5 Pcs/Pkt
59	Glass Slab for dental cement mixing	Dimension Size L-120 MM,H-7 MM and 78 MM	Each
60	Spoon Excavator	Sharp Stainless Steel High Quality Size No #1, #1S	Each
61	Erich Arch Bar	It should be made up from high quality stainless steel	Each
List of Instrument			
1	Lower Molar Forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
2	Lower Pre Molar Forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
3	Lower anterior Forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
4	Upper Molar Forceps left	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
5	Upper Molar Forceps right	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
6	Upper Pre-molar forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
7	Upper Anterior Forceps dental	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
8	Upper root Forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
9	Upper cow horn forceps left	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
10	Upper cow horn forceps Right	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
11	Cryer elevators	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Pair
12	Coupland Elevator	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
13	Warwick james elevator	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
14	Periosteal Elevator	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
15	Straight Elevator	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each

16	Curved Scissor	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
17	Mosquito Artery forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
18	Needle Holder	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
19	Curved Needle for Dental Use	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Pkt
20	B.P.Handle for No 15 Blade	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
21	Surgical curette	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
22	Mouth Mirror	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
23	Dental Explorer	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
24	Tweezers for dental use	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
25	Bone rongeur for dental use	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
26	Bone file	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
27	Surgical Burs (702)	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
28	Upper 3rd Molar Forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
29	Pedo extraction forceps	High quality stainless-steel, anti-magnetic, rust resistant, acid-resistant and should comply to above mentioned instrument product quality standards.	Each
30	Suction Tip Dental		Pkt
31	Instrument Sterilizer	<p>Approx. Size (L X B X H): 300x150x125mm (Size 12" X 6" X 5")- 1 KW Load Made of seamless stainless steel-304 Grade Automatic tray lifting system Handle on side with insulated grip Thermostatically controlled With water outlet tap 220-240 Volts, single phase</p>	Each


Chief District Medical & Public Health Officer,
Bargarh

LIST OF ITEMS QUOTED

(To be attached with technical bid)

Sl.No	Item Sl.No as per tender	Specification	Name of the Manufacturer	Product Certification (if yes provide a copy)	ISO certified company (if yes provide a copy.	Remarks(Please mention if there is any deviation to quoted specification
1						
2						

PRICE SCHEDULE

ANEXURE-II

Sl.No	Item Sl.No.as per the tender	Name of the Item	Make/Model	Name of the Manufacturer	Rate in (Rs.) for Each Nos/ Packet/ Set as per Annexure -I	Taxes %	Total Unit Cost Including Taxes

N.B: Please provide both softcopy and hardcopy in the price bid

!

DECLARATION FORM

I / We _____ having My / our _____ office at

_____ do declare that I / We have carefully read all the terms & conditions of tender of the C.D.M. & P.H.O,Bargarh, Orissa for the supply of EIF. I will abide with all the terms & conditions set forth in the tender paper Reference no. _____.

I/We do hereby declare I/We have not been de-recognized / black listed by any State Govt. / Union Territory / Govt. of India / Govt. organization / Govt. Health Institutions for supply of Not of Standard Quality(NSQ) items / part-supply / non-supply.

I/We agree that the Tender Inviting Authority can forfeit the Earnest Money Deposit and or Security Deposit and blacklist me/us for a period of 2 years if, any information furnished by us proved to be false at the time of inspection / verification and not complying with the Tender terms & conditions.

I / We further declare that I / We possess valid manufacturing license / authorised distributor license bearing No. _____ Valid upto _____ I / We _____ do hereby declare that I / we will supply the _____(List of Items- attached separate sheet) as per the terms, conditions & specifications of the tender document. I / we further declare that my / our EMD / performance security will be forfeited if I / we fail to supply any item after getting order from the purchaser.

Signature of the bidder :

Date :

Name & Address of the Firm: Affidavit before Executive Magistrate / Notary Public.

ANNUAL TURN OVER STATEMENT

The Annual Turnover of M/s _____
_____ who is a manufacturing unit / wholesaler / distributor for the
last three years are given below and certified that the statement is true and correct.

<i>Sl. No.</i>	<i>Year</i>	<i>Turnover in Lakhs / Crores (Rs)</i>
1.	2021 - 2022	-
2.	2022 – 2023	-
3.	2023 – 2024	-

Date: _____ Signature of Auditor/
Place: _____ Chartered Accountant
(Name in Capital)
Registration No.
Seal

NB:

1. This certificate should be supported by figures in PL Account & Income Tax Return.

MANUFACTURER’S AUTHORISATION FORMAT

(in original in case bidder is authorized by manufacturer)

To

The CDM &PHO
Bargarh

Sub: LETTER OF AUTHORISATION.

Ref: Tender Reference No.

Respected Sir,

We _____ who are established and reputed manufacturer’s of _____ (name and description of items offered) having factories at _____ (Address of Factory) do hereby authorize M/s _____

_____ (Address of Factory) do hereby authorize M/s _____

_____ (Name and address of Distributor / Agent) to submit a bid and sign the contract with you against the above referred tender.

We also extend our full quality assurance for the items quoted by M/s _____ as per the terms and conditions in your tender under reference above.

Yours faithfully,

Full Name of the designated person

(Signature with seal)

Contact Number:

Email ID:

Note: This letter of authority should be on the letter head of the manufacturer and should be signed by a person competent and having the power of attorney to bind the manufacturer. It should be included in the bid submitted by the bidder if the bidder is not the manufacturer.